



State of Wisconsin  
Governor Tony Evers

**Department of Agriculture, Trade and Consumer Protection**

Wisconsin Weights and Measures Laboratory

**Calibration Certificate**

*for calibration work performed for:*

**FOX VALLEY INDUSTRIAL SCALE, INC.**

109 FORD DR STE D  
NEW LENOX, IL 60451-3669  
(815) 463-1209

Date Received:	March 24, 2023	State Test No.:	W23-108
Date of Calibration:	March 24, 2023		
Date Issued:	March 27, 2023		
Date Due:	March 24, 2025		

**Uncertainty Statement**

For the mass standards used in this calibration, some uncertainty components were assessed through a Type A evaluation, the method for assessing uncertainty by a statistical analysis of measured quantity values obtained under defined measurement conditions. In addition, other components were assessed from a Type B evaluation of standard uncertainty, based on scientific judgement using all of the relevant information available. The combined standard uncertainty was multiplied by a statistically determined coverage factor to provide an expanded uncertainty. The expanded uncertainty defines an interval having a level of confidence of approximately 95 percent, assuming normal distribution. The expanded uncertainty presented in this report is consistent with the ISO/IEC Guide to the Expression of Uncertainty in Measurement using the Root Sum Squares method (JCGM 100:2008).

**Traceability Statement**

The standards used by the Wisconsin State laboratory demonstrate an unbroken traceable chain to the International System of Units (SI) through the National Institute of Standards and Technology (NIST) and are part of a comprehensive measurement assurance program for ensuring continued accuracy and measurement traceability within the level of uncertainty reported by this laboratory. The laboratory maintains documented calibration intervals and uses documented procedures, all under the performance of trained personnel who demonstrate suitable measurement assurance for the information listed in this calibration certificate. The laboratory test number identified above is the unique test number to be used in referencing measurement traceability for the artifacts identified in this certificate. The State Standards are traceable to the SI unit for mass, the kilogram.

**Conformity Statement**

These results relate only to the items calibrated in this certificate. Field standards and weight carts are calibrated based on guidance described in NIST Handbook 105-1 (2019) and NIST Handbook 105-8 (2019), respectively, using NISTIR 6969: Selected Laboratory Measurement Practices and Procedures to Support Basic Mass Calibrations (2019). Field standards calibrated to NIST Class F, ASTM 5, and ASTM 6 tolerances are usable for testing class III, III L, and IIII weighing devices, following NIST Handbook 44, Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices. Field standards calibrated to NIST Class F, ASTM 5, or ASTM 6 tolerances are not suitable for testing class I and class II weighing devices, which must be tested with field standards of higher precision than NIST Class F, ASTM 5, or ASTM 6. Weights calibrated to ASTM 7 tolerances by this laboratory cannot be used for testing commercial weighing devices. Field standards calibrated to ASTM Standard Specification E617-18 are not checked for density [Stainless steel weights are assumed 8.0 grams per cubic centimeter], or for magnetism.

**Decision Rule**

All calibrated weights and weight carts that are determined to have a mass correction such that:  $|\text{Correction}| > (\text{Tolerance} - \text{Uncertainty})$  are considered to have failed to meet the applicable tolerance. It is the decision rule of the Wisconsin State laboratory that all calibrated weights and weight carts that are determined to have a mass correction such that:  $|\text{Correction}| > (0.95 * \text{Tolerance} - \text{Uncertainty})$  will be adjusted to be closer to zero mass correction, even if the mass correction of the weights and weight carts originally met the applicable tolerance. Customers may request exceptions to this decision rule.

The following standard(s) were used: 1000 lb: 392, 25 lb: 25LBCS

*This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.*

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director



State of Wisconsin  
Governor Tony Evers

**Department of Agriculture, Trade and Consumer Protection**

Wisconsin Weights and Measures Laboratory

**Calibration Certificate**

Date Received: March 24, 2023  
Date of Calibration: March 24, 2023  
Date Issued: March 27, 2023  
Date Due: March 24, 2025

State Test No.: W23-108  
Item(s) Submitted: Cast Weight  
Manufacturer: Rice Lake  
Condition: Good, Acceptable for Calibration  
Tolerance Class: NIST HB 105-1 (1990), Class F

Customer: FOX VALLEY INDUSTRIAL SCALE, INC.  
Address: 109 FORD DR STE D  
NEW LENOX, IL 60451-3669  
Contact: WENDY ECKBERG  
Phone: (815) 463-1209

Balance ID#: 8&10  
Procedure Used: NISTIR 6969 (2019), SOP 8  
Temperature: 20.7 °C  
Relative Humidity: 46.0 %  
Pressure: 739.0 mmHg

Nominal Mass	Mass Unit	Serial No.	Conventional Mass Correction (mg)		NIST HB 105-1 (1990), Class F		Uncertainty (mg)	Coverage Factor (k)
			As Found	As Left	As Found	As Left		
1000	lb	108	3,100	3,100	Pass	Pass	5,900	2.00
1000	lb	106	2,500	2,500	Pass	Pass	5,900	2.00
1000	lb	103	5,000	5,000	Pass	Pass	5,900	2.00
1000	lb	107	400	400	Pass	Pass	5,900	2.00
1000	lb	105	2,800	2,800	Pass	Pass	5,900	2.00
1000	lb	110	6,700	6,700	Pass	Pass	5,900	2.00
1000	lb	102	4,200	4,200	Pass	Pass	5,900	2.00
1000	lb	100	5,800	5,800	Pass	Pass	5,900	2.00
1000	lb	111	2,400	2,400	Pass	Pass	5,900	2.00
1000	lb	101	8,600	8,600	Pass	Pass	5,900	2.00
1000	lb	109	5,700	5,700	Pass	Pass	5,900	2.00
1000	lb	104	300	300	Pass	Pass	5,900	2.00
25	lb	3	140	140	Pass	Pass	140	2.01
25	lb	00	-260	-260	Pass	Pass	140	2.01

The following standard(s) were used: 1000 lb: 392, 25 lb: 25LBCS

*This report may not be reproduced, except in full, without the written approval of the Lead Metrologist or Laboratory Director.*

Paul Masterson, Lead Metrologist

Justin Lien, Laboratory Director